

Certificate of Analysis

| | |
|-----------------------|--|
| CATALOG NUMBER | CUV-NA |
| LOT NUMBER | A-1120 |
| SPECIFICATION | Absorbance Target Value (1cm): 0.675 - 0.825 |
| THIS LOT VALUE | .747 |

Product Description: CUV-NA is a secondary reference material for verifying performance of DeNovix® Spectrophotometers. The solution is gravimetrically prepared using reagent-grade nicotinic acid and spectroscopy-grade water. NOTE: Each ampoule of CUV-NA is intended for single use within 20 minutes of opening.

QC Test: CUV-NA was measured on an Agilent 8453 spectrophotometer using spectroscopy-grade water as a reference. The 261nm absorbance was measured bi-chromatically (340 nm baseline reference) with a 1cm pathlength absorbance per Beer's law.

Additional Info: An Agilent 8453 spectrophotometer is annually certified using traceable standard solutions for absorbance accuracy, wavelength accuracy, stray light, noise, and spectral resolution by an operational qualification procedure performed by the vendor.

I certify that the above information is accurate and true.



David Ash, Laboratory Manager